

UKAS LoadTesting

DECLARATION OF LOAD TEST RESULTS

PRODUCT SUPPLIER QUINTAL HEALTHCARE LTD

SUPPLIER ADDRESS Westminster Business Centre
10 Great North Way
Nether Poppleton
York
YO26 6RB
United Kingdom

Declares that the Product described below

PRODUCT NAME: Q2 Motility Board
COLOUR: Blue
BATCH: Q2-0000

was load tested at the UKAS approved Independent Test Laboratory below

COMPANY: CERAM Research
ADDRESS: Queens Road
Penkhul
Stoke-on-Trent
ST4 7LQ
England

TEST ENGINEER: Mr David Dix
TEST VENUE Queens Road Laboratories-Stoke on Trent
TEST DATE 30th June 2005

LOAD TESTING PROCEDURE:

The Q2 Motility Board was load tested by applying a load hydraulically to a soft bottomed Uniformly Distributed load. The actual UDL sizes were determined from standard anthropometry data and the actual loads applied were measured by means of a load cell. Board deflection was monitored with deflection transducers.

LOAD TEST RESULTS

Samples were tested in varying support configurations with different UDLs. In all cases the Q2 Motility Board withstood loads significantly in excess of its authorised Working Load Limit of **285 kg** with acceptable deflection.

Signed by on behalf of the Manufacturer



Colin Quickmire

Position: Director and CE Responsible Person

Signed at: York England on July 1st 2005

Full details of the test procedure and results are available for viewing, if required